





LipiView® II – Leading Innovation in Lipid Layer Thickness Measurement and Meibomian Gland Imaging.



Exmple OS Image – Average Lipid Layer Thickness 100+nm

Nanometer Dynamic Lipid Layer Thickness


Features patented technology analyzing over one billion data points to isolate lipid layer of the tear film, measuring thickness to sub-micron levels.



Example Partial Blink

Blink Analysis

Evaluates dynamic response of the lipids to blinking, while proprietary algorithms measure the extent of lid closure during each blink.



DMI – Adaptive Transillumination

High-definition Gland Imaging

Utilizes Dynamic Meibomian Imaging™ (DMI) to distinctively image meibomian gland structure in high definition.

Please see indication and important safety information on back page.

PP2018TS4252



INDICATIONS AND IMPORTANT SAFETY INFORMATION for LIPIVIEW® II Ocular Surface Interferometer

Rx ONLY

INDICATIONS

The LipiView® II Ocular Surface Interferometer is an ophthalmic imaging device that is intended for use by a physician in adult patients to capture, archive, manipulate and store digital images of:

- Specular (interferometric) observations of the tear film. Using these images, LipiView® II measures the absolute thickness of the tear film lipid layer.
- Meibomian glands under nearinfrared (NIR) illumination.
- The ocular surface and eyelids under white illumination.

CONTRAINDICATIONS

Contraindications are conditions in which the device should not be used because

the risk of use clearly outweighs any benefit. No contraindications have been identified for LipiView® II.

PRECAUTIONS

The following patient conditions may affect the interferometry assessment of a patient's tear film using LipiView® II:

- Use of ophthalmic drops such as artificial tear lubricants, ointments, and medications. Advise patients not to instill oil-based ophthalmic drops (e.g., Soothe®, Restasis®, Systane Balance®) for at least 12 hours prior to device use and not to instill ointments for at least 24 hours prior to device use. Wait at least four (4) hours after the instillation of all other ophthalmic drops prior to device use.
- Soft or rigid contact lens wear. Advise patients to remove contact lenses at least four hours prior to device use.

- Use of oil-based facial cosmetics around the eye.

Eye rubbing.

- Recent swimming in a chlorinated pool. Advise patients to not to swim for at least 12 hours prior to device use.
- Any ocular surface condition that affects the stability of the tear film. These conditions include disease, dystrophy, trauma, scarring, surgery, or abnormality.

ADVERSE EFFECTS

There are no known or anticipated adverse effects associated with use of this device.

ATTENTION

Reference the *LipiView® II Ocular Surface Interferometer Instructions for Use* for a complete listing of indications, warnings, and precautions.

LIPIVIEW® II SPECIFICATIONS

Weight	52 lbs
Dimension	17"W x 31"D x 18"H